July 15, 2015 Products Committee Meeting: Attending: Neal Burkhardt, Dr. Collins, Chris Grove, Scott Hansen, Deb Kuhlman, Korey Lopes, Mary Lewis, Roseni Roche, Tosha Whitley Absent: Connie Anderson, Joyce Morimoto, Emiliah Sithole, Ed Stauffer,

ITEM	Discussion	Action	Completion	Follow-up
Safety syringe Vanish Point 1ml	Have looked at this device in the past; \$35.57/box 100	Excellent device Look to add syringe with 27g needle	Closed	Change to vendor
Vanish Point syringes 3ml	Excellent device not seen as a need as there haven't been frequent needle stick with this size syringe	No change with inventory	Close	Reopen if needed
<u>Critline</u>	NKC to look at Critline product	Demo held – revised product due to be released this summer	HOLD	Hold until revised model available Critline #4
<u>Citrisate</u>	NKC to look at Citrisate mixing	NKC to look at trial long term change	HOLD	Reopen January 2016
Dialyzer?	Off Menu dialyzer	Can F250 be used routinely? <b>NO</b>	Open	Administration to provide directions Plan to limit menu
	Dr. Collins:	F180 \$7.20; F200 \$8.92; F250 \$9.65 (29 patients now using)		
Chlorhexidine <3% Correction to May Meeting minutes	2% Chloraprep Single swabstick \$0.42 Or 2% ChloraPrep One-Step \$0.51	NKC to continue with current product (\$0.14) and add the single swab stick as an alternative strength for patients with sensitivities	Closed	Correction to May's meeting minutes
ARJO slings	Sling safety, product life expectancy, cleaning	Life approximately 2yrs, cleaning with disinfectant spray and if soiled laundry as a medical device.	Closed	Tosh review linen contract; Procedure edits to be added
Fistula needle with retracting sharp	Many needle sticks due to poor use of safety device & difficulties for staff with short fistula segments	No know new products to review	Closed	Reopen as new fistula needles are identified
1k/2.5ca jug	\$11.82/case; cost effective eliminate Granuflo acid mixed & dumped	Roll out with FSS group	Closed	To review at their next meeting
Sit-to-Stand	Assists patient movement to stand from a chair May be useful with getting patients to/from bathroom	Further discussion with the Unit Managers – Tuesday, July 21	OPEN	Discuss at September Manager meeting
Dialysis chair	If Trendelenburg is not a requirement revisit dialysis chairs	Look at patient comfort; look at chair maintenance	OPEN	Discuss with FSS chair repair/maintenance
Myron-L syringe	New policy coming	Single use 60 cc syringe	Closed	Mary to send policy change to clinical units

Next Meeting: September16, 2015 10:30am Melange Conference Room